



CERTIFICATE OF ANALYSIS

Product: Glycerine Swabsticks
Description: Lemon flavour
Product reference: 1115
Lot number: 25102021
Expiry date: 10/2026

| Test | Specification | Results |
|----------------------------------|---------------|---------|
| Swab size (length) | 100 ± 10 mm | Pass |
| Swab size (cotton diameter) | 10 ± 1 mm | Pass |
| Sealing | No leakage | Pass |
| Lemon glycerine solution content | 4.5 ± 0.4 ml | Pass |

This testing has been carried out by a third-party company.

Claire Sykes
Senior Quality Associate
24-Aug-2022

